

## LCR Hi-TESTER (HIOKI 3532-50)

In charge: Dr. S. Manivannan



### Specifications:

1. Measurement parameters:  $|Z|$ , R, C,  $\tan \delta$
2. Test frequency range: 42 Hz to 5 MHz
3. Input Z: 50  $\Omega$
4. AC Test Signal level: 10 mV to 5 V
5. Measurement range: 10  $\mu$ A to 100 mA rms (DC to 1MHz)  
50  $\mu$ A to 20 mA rms (1MHz to 5MHz)
6. DC bias: External DC bias  $\pm$  40 V max.
7. Operating: 240 V ( $\pm$  10%) AC, 50/60 Hz.
8. Sample type: Solid samples