

National Institute of Technology, Tiruchirappalli-15

Tender for procurement of Contact Angel Meter

Tender Number: NITT/F/No.UG 010 / Plan 2014-15/ MEC dated 16/11/2015

Minutes of the pre bid meeting held on 24/11/2015 at the office of the Dean
(R&C) at 3.30 P.M

Members present

Bidder representative

Ashish Gusain, Orbit Research Associate Private Limited, Delhi

NITT representative

Dr.S.Suresh, Purchase Initiator

G.Soundara rajan, Deputy Registrar

At the outset Deputy Registrar asked the bidder representative that whether they require any clarification with reference to technical or commercial conditions stipulated in the tender.

In response, M/s Orbit Research Associate Private Limited, Delhi stated that 6.5 X is the standard and universally accepted specification for zoom capability of high speed camera and as such requested NITT to review the Camera Zoom Capability.

After due deliberations, the purchase committee members of NITT agreed to revise the zoom capability capability as 6.5 X, as against 10 x stipulated in the tender.

There is no change in the other technical specifications.



Dr. S. SURESH, M.Tech., Ph.D.
Assistant Professor
Department of Mechanical Engineering,
National Institute of Technology
Tiruchirappalli - 620 015, Tamil Nadu, India