



Sub: Tender for Supply and Installation of Ultrapure Water Purification system.

Ref: NITT/F.NO.028/CAP.EXP.35/2019-20/CHY dated: 13.05.2020

Based on the purchase committee discussion the following corrigendum points were published:

TECHNICAL SPECIFICATIONS		
PUBLISHED	AMENDED TECHNICAL SPECIFICATIONS AND IT SHOULD BE READ AS	
Ultrapure Water Purification system	Tap Water feed (Feed water conductivity : <2000 μS/cm) Integrated table top Water Purification System	Tap Water feed (Feed water conductivity : <2000 μS/cm) Integrated table top or table top Water Purification System
	Stage I - Reverse Osmosis with de-ionization:	Stage I - Reverse Osmosis with de-ionization or Electro de-ionization
	Mixed bed resin de-ionization module for reduction of the inlet conductivity.	Mixed bed resin Electro de-ionization or de-ionization module for reduction of the inlet conductivity.
	Production Rate : 20 Liters per hour or above	Production Rate : 10 Liters per hour or above
Last Date	<u>The Last date for online bid submission end date is extended up-to 15-June-2020 (03.00PM) and Technical Bid opening will be on 16-June-2020 (03.00PM)</u>	

Hence, the bidders are requested to quote the same in PDF as well as in Excel format and the same to be uploaded in the E-Tender portal.

The rest of the technical specifications of all other items and the terms and conditions of the aforementioned E-tender remain unaltered.

Tender Inviting Authority