



CORRIGENDUM

TENDER NOTIFICATION NO.: NITT. No.004/HEFA/2019-20/SIEM/CoDE DATED 24/10/2019

E-Tender ID : 2019_NITT_516710_1



for the Upgradation of existing pico-second laser to femtosecond laser

BASED ON MINUTES OF PRE BID MEETING HELD ON 14/11/2019 11.00A.M. AT NATIONAL INSTITUTE OF TECHNOLOGY, TIRUCHIRAPPALLI – 620 015
THIS CORRIGENDUM HAS BEEN PUBLISHED.

Based on the discussion and, the committee members finally decided to amend the following points:

Sl.No.	Tender Document Details	Bidders raised clarifications in Published document	The following points to be read as follows against the published document under NITT's Specification				
1	Part 2 Supply Requirements Section VI. Schedule of Requirements	Page 29 under Details of Laser Source <table border="1"> <tr> <td>Wavelength (nm)</td> <td>1030 ± 5</td> </tr> </table>	Wavelength (nm)	1030 ± 5	Page 29 under Details of Laser Source is amended and the bidders should read as <table border="1"> <tr> <td>Wavelength (nm)</td> <td>1030 ± 10</td> </tr> </table>	Wavelength (nm)	1030 ± 10
Wavelength (nm)		1030 ± 5					
Wavelength (nm)		1030 ± 10					
2	Page 29 under Details of Laser Source <table border="1"> <tr> <td>Pulse Width</td> <td>Variable in the range of 350 fs (or lower) to 5 ps (or higher)</td> </tr> </table>	Pulse Width	Variable in the range of 350 fs (or lower) to 5 ps (or higher)	Page 29 under Details of Laser Source is amended and the bidders should read as <table border="1"> <tr> <td>Pulse Width</td> <td>"Variable in the range of 400 fs (or lower) to 5 ps (or higher)."</td> </tr> </table>	Pulse Width	"Variable in the range of 400 fs (or lower) to 5 ps (or higher)."	
Pulse Width	Variable in the range of 350 fs (or lower) to 5 ps (or higher)						
Pulse Width	"Variable in the range of 400 fs (or lower) to 5 ps (or higher)."						
3	Page 30 under Workstation	Page 30 under Workstation <u>Additional one point added along with the existing, the bidders should note the same</u> ✓ "Samples should be processed with spot size of 100 µm or lower".					

The photographs of the existing laser system are appended in the next page for reference.



Tender Inviting Authority